

CALIBRATION CHART

2B-12 INCREMENTAL DISPLACEMENT = 7.24 μ L/THOU

2B-14 INCREMENTAL DISPLACEMENT = 9.85 μ L/THOU

2B-17 INCREMENTAL DISPLACEMENT = 14.53 μ L/THOU

2B-20 INCREMENTAL DISPLACEMENT = 20.11 μ L/THOU

