

CALIBRATION CHART

5H3-100 ML INCREMENTAL DISPLACEMENT = 0.034 ML/THOU
5H3-250 ML INCREMENTAL DISPLACEMENT = 0.097 ML/THOU
5H3-500 ML INCREMENTAL DISPLACEMENT = 0.180 ML/THOU
5H3-1000 ML INCREMENTAL DISPLACEMENT = 0.337 ML/THOU

